



Rinse mouth with clean water, 30 minutes prior to collection. Do not eat, drink, smoke, or chew gum after rinsing.



Spit into the funnel until the amount of saliva reaches the "Saliva" line.



Holding the tube upright, unscrew the funnel from the tube gently. This releases the saliva into the receptacle below, containing a clear fluid.



Close the tube using the blue cap, tightly. Shake the capped tube for 5 seconds.



Place the tube into the Sample Return Bag and seal. Discard the funnel.

